

Hypoglycemia Prediction with Subject-Specific Recursive Time-Series Models

Meriyan Eren-Oruklu, M.E.,¹ Ali Cinar, Ph.D.,¹ and Laretta Quinn, Ph.D., R.N.²

Abstract

Background:

Avoiding hypoglycemia while keeping glucose within the narrow normoglycemic range (70–120 mg/dl) is a major challenge for patients with type 1 diabetes. Continuous glucose monitors can provide hypoglycemic alarms when the measured glucose decreases below a threshold. However, a better approach is to provide an early alarm that predicts a hypoglycemic episode before it occurs, allowing enough time for the patient to take the necessary precaution to avoid hypoglycemia.

Methods:

We have previously proposed subject-specific recursive models for the prediction of future glucose concentrations and evaluated their prediction performance. In this work, our objective was to evaluate this algorithm further to predict hypoglycemia and provide early hypoglycemic alarms. Three different methods were proposed for alarm decision, where (A) absolute predicted glucose values, (B) cumulative-sum (CUSUM) control chart, and (C) exponentially weighted moving-average (EWMA) control chart were used. Each method was validated using data from the Diabetes Research in Children Network, which consist of measurements from a continuous glucose sensor during an insulin-induced hypoglycemia. Reference serum glucose measurements were used to determine the sensitivity to predict hypoglycemia and the false alarm rate.

Results:

With the hypoglycemic threshold set to 60 mg/dl, sensitivity of 89, 87.5, and 89% and specificity of 67, 74, and 78% were reported for methods A, B, and C, respectively. Mean values for time to detection were 30 ± 5.51 (A), 25.8 ± 6.46 (B), and 27.7 ± 5.32 (C) minutes.

Conclusions:

Compared to the absolute value method, both CUSUM and EWMA methods behaved more conservatively before raising an alarm (reduced time to detection), which significantly decreased the false alarm rate and increased the specificity.

J Diabetes Sci Technol 2010;4(1):25-33

Author Affiliations: ¹Department of Chemical and Biological Engineering, Illinois Institute of Technology, Chicago, Illinois; and ²College of Nursing, University of Illinois at Chicago, Chicago, Illinois

Abbreviations: (ARMA) autoregressive moving-average model, (CG-EGA) continuous glucose–error grid analysis, (CGM) continuous glucose monitoring, (CGMS) Continuous Glucose Monitoring System, (CUSUM) cumulative-sum control chart, (DirecNet) Diabetes Research in Children Network, (EWMA) exponentially weighted moving-average control chart, (LCL) lower control limit, (WRLS) weighted recursive least squares

Keywords: CUSUM control chart, early hypoglycemic alarms, EWMA control chart, glucose predicting models, time-series methods

Corresponding Author: Ali Cinar, Ph.D., Department of Chemical and Biological Engineering, Illinois Institute of Technology, 10 W. 33rd Street Perlstein Hall, Suite 127, Chicago, IL 60616; email address cinar@iit.edu